

Coagulation							
External Name	Abbrev	Sex	Age From	Age To	Result Checking Type	Numeric Range (N)	Units
Antithrombin III Activity Assay	ATIII	~	0	1 Months	Reference Range	60-89	%
Antithrombin III Activity Assay	ATIII	~	1 Months	1 Years	Reference Range	72-134	%
Antithrombin III Activity Assay	ATIII	~	1 Years	6 Years	Reference Range	101-131	%
Antithrombin III Activity Assay	ATIII	~	6 Years	11 Years	Reference Range	95-134	%
Antithrombin III Activity Assay	ATIII	~	11 Years	16 Years	Reference Range	96-126	%
Antithrombin III Activity Assay	ATIII	~	16 Years	150 Years	Reference Range	83-128	%
Chromogenic Factor X Level	FactorXChr	~	~ ~	~ ~	Reference Range	>=15	%
Chromogenic Factor X Level	FactorXChr	~	~ ~	~ ~	Critical	<15	%
D-Dimer (FEU)	Dimer-FEU	~	~ ~	~ ~	Reference Range	<=574	ng/mL
Factor IX Assay	Factor IX	~	0	1 Months	Reference Range	44.0-97.0	%
Factor IX Assay	Factor IX	~	1 Months	1 Years	Reference Range	43.0-121.0	%
Factor IX Assay	Factor IX	~	1 Years	6 Years	Reference Range	44.0-127.0	%
Factor IX Assay	Factor IX	~	6 Years	11 Years	Reference Range	48.0-145.0	%
Factor IX Assay	Factor IX	~	11 Years	16 Years	Reference Range	64.0-216.0	%
Factor IX Assay	Factor IX	~	16 Years	150 Years	Reference Range	65.0-150.0	%
Factor V Assay	Factor V	~	0	1 Months	Reference Range	92-154	%
Factor V Assay	Factor V	~	1 Months	1 Years	Reference Range	94-141	%
Factor V Assay	Factor V	~	1 Years	6 Years	Reference Range	67-127	%
Factor V Assay	Factor V	~	6 Years	11 Years	Reference Range	56-141	%
Factor V Assay	Factor V	~	11 Years	16 Years	Reference Range	67-141	%
Factor V Assay	Factor V	~	16 Years	150 Years	Reference Range	62-139	%
Factor VII Assay	Factor VII	~	0	1 Months	Reference Range	67-107	%
Factor VII Assay	Factor VII	~	1 Months	1 Years	Reference Range	83-160	%
Factor VII Assay	Factor VII	~	1 Years	6 Years	Reference Range	72-150	%
Factor VII Assay	Factor VII	~	6 Years	11 Years	Reference Range	70-156	%
Factor VII Assay	Factor VII	~	11 Years	16 Years	Reference Range	69-200	%
Factor VII Assay	Factor VII	~	16 Years	150 Years	Reference Range	63-174	%
Factor VIII Assay	FactorVIII	~	0	1 Months	Reference Range	83.0-274.0	%
Factor VIII Assay	FactorVIII	~	1 Months	1 Years	Reference Range	54.0-145.0	%
Factor VIII Assay	FactorVIII	~	1 Years	6 Years	Reference Range	36.0-185.0	%
Factor VIII Assay	FactorVIII	~	6 Years	11 Years	Reference Range	52.0-182.0	%
Factor VIII Assay	FactorVIII	~	11 Years	16 Years	Reference Range	59.0-200.0	%
Factor VIII Assay	FactorVIII	~	16 Years	150 Years	Reference Range	50.0-185.0	%
Fibrinogen	Fibrinogen	~	~ ~	~ ~	Reference Range	200-393	mg/dL
Fondaparinux Level	Fonda	~	~ ~	~ ~	Reference Range	<=1.30	mg/L
Fondaparinux Level	Fonda	~	~ ~	~ ~	Critical	>1.30	mg/L
Heparin Level	Heparin	~	~ ~	~ ~	Reference Range	<=2.00	unit/mL
Heparin Level	Heparin	~	~ ~	~ ~	Critical	>2.00	unit/mL
International Normalization Ratio	INR	~	~ ~	~ ~	Reference Range	<=5.0	
International Normalization Ratio	INR	~	~ ~	~ ~	Critical	>5.0	
MCS Protocol Plasma HGB	MCS Plhgb	~	~ ~	~ ~	Reference Range	<30	mg/dL
Partial Thromboplastin Time	PTT Irrabb	~	0	30 Days	Reference Range	27.0-39.2	second(s)

Coagulation							
External Name	Abbrev	Sex	Age From	Age To	Result Checking Type	Numeric Range (N)	Units
Partial Thromboplastin Time	PTT Irrabb	~	0	30 Days	Critical	>249.0	second(s)
Partial Thromboplastin Time	PTT Irrabb	~	30 Days	12 Years	Reference Range	27.0-39.2	second(s)
Partial Thromboplastin Time	PTT Irrabb	~	30 Days	12 Years	Critical	>249.0	second(s)
Partial Thromboplastin Time	PTT Irrabb	~	12 Years	150 Years	Reference Range	27.0-39.2	second(s)
Partial Thromboplastin Time	PTT Irrabb	~	12 Years	150 Years	Critical	>249.0	second(s)
PFA-ADP	PFA-ADP	~	~ ~	~ ~	Reference Range	62-117	second(s)
PFA-ADP INTERP		~	~ ~	~ ~	Reference Range		
PFA-EPI	PFA-EPI	~	~ ~	~ ~	Reference Range	85-185	second(s)
PFA-EPI INTERP		~	~ ~	~ ~	Reference Range		
Prothrombin Time	PT	~	0	30 Days	Reference Range	9.8-12.9	second(s)
Prothrombin Time	PT	~	30 Days	1 Years	Reference Range	9.8-12.9	second(s)
Prothrombin Time	PT	~	1 Years	150 Years	Reference Range	9.8-12.9	second(s)